

FEB 25 2011  
3/18/11

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>AW Anderson</u>	ID# <u>2467</u> Date <u>2/23/11</u>
A Agency <u>Seldovia Police Dept.</u> Phone # _____	
Instrument Location <u>Seldovia Police Dept.</u>	
DataMaster cdm S/N <u>130180</u>	
B Alco S/N <u>X301530</u> Target Value <u>.078</u> High Pressure <u>500 PSI</u>	
Alco Test Values	
<u>.078</u> 1st Alco	<u>.078</u> 2nd Alco
Signature <u>AW Anderson</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of MAR, 2011

Daniel L. Nicholas  
(Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130180

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130180

FEBRUARY 23, 2011

OPERATOR'S NAME:

ANDERSON AM

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIXTY DAY VERIFICATION

SUBJECT'S FIRST NAME/MI :

N/A

D.L. #: 1234567890

DEPT/AGENCY: SOU1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: K301530

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 31.83 in  
ALCO TARGET .000 11:53  
BLANK TEST .000 11:54  
INTERNAL STANDARD VERIFIED 11:54  
ALCO TO 31.83 in .078 11:55  
BLANK TEST .000 11:55  
SUBJECT SAMPLE .000 11:56  
BLANK TEST .000 11:57  
ALCO TO 31.83 in .078 11:57  
BLANK TEST .000 11:58

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130180

FEBRUARY 23, 2011

TIME 11:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 41c

BAROMETER: 31.83 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abdefghijklmnop  
qrstuvwxyz{|}~